

3034T4



# Sicherheitsschrank Typ 30 - EN 14470-1 - Hoch 1 Tür - 4 Schubladen

## ÜBEREINSTIMMUNG

EN 14470-1, FM 6050

## ADVANTAGES

- 30-minute fire resistant and made in France
- Storage of flammables

## SONDERN AUCH...

- Thermo-expanding seals of the doors
- Holes for the connection to the ventilation
- Normalized labelling provided with the equipment
- Adjustable leveling feet for enhance stability
- Geeignet für den Transport mit Hubwagen
- Tested and approved according to FM 6050
- Drawers with automatic closure in case of a fire



## TECHNISCHE DATEN

Referenz  
Übereinstimmung  
Baustoffe der Sicherheitsschränke

**3034T4**  
**EN 14470-1, FM 6050**  
**Melamine**

Türfarbe  
Anzahl der Türen  
Art der Türen  
Scharniere

**hellgrau RAL 7035**  
**2**  
**flügeltüren**  
**rechts**

Verriegelung der Türen  
Versandabmessungen HxBxT (mm)  
Belüftung

Durchmesser der Öffnungen (mm)  
Anzahl der Abteile / Fächer  
Codenummer

**selbstschließend**  
**2180 x 1200 x 800**  
**zuluftöffnung und abluftöffnung**  
**auf dem dach**  
**100**  
**1**  
**94036000**

## COMPLEMENTARY ACCESSORIES - VENTILATION AND FILTRATION BOXES

- Filteraufsätze mit filter für organische Dämpfe (Réf : CDF-A)
- Lüftungskasten (Réf : CDV-A)
- Filteraufsätze (ohne Filter) - NF X 15-211 (Réf : H50C)
- Anschlüsse für Filteraufsätze (Réf : KRC)

| Referenz | Außere Abmessungen HxBxT (mm) | Innere Abmessungen HxBxT (mm) | Lagerkapazität (L) | Auffangwanne | Regale | Tragfähigkeit | Gewicht (Kg) |
|----------|-------------------------------|-------------------------------|--------------------|--------------|--------|---------------|--------------|
| 3034T4   | 1980 x 605 x 550              | 1581 x 491 x 419              | 130                |              | 0      | 30            | 141.00       |



**TRIONYX**  
1<sup>er</sup> FABRICANT FRANÇAIS  
Armoires de sécurité pour produits dangereux

138 avenue Henri BARBUSSE 92140 Clamart (France)  
Tél. 01 46 45 80 00 - Fax : 01 46 45 85 98  
www.trionyx.fr - trionyx[at]trionyx.fr